

## Certificate of Analysis

**Product Description:** NATtrol™ Adenovirus Type 40 Stock

**Strain:** Dugan

**Catalog Number:** NATADV40-ST

**Lot Number:** 329801

**Expiration Date:** 2023-08-01

**Storage Temperature:** 2-8°C

**Safety Precautions:** Use Universal Precautions.

Test	Specification	Ct Value	Result
Nucleic acid detection by PCR	22.0-25.0	22.2	Pass
Inactivation verification*	Inactive		Pass

\*Endpoint dilution assays were used to evaluate inactivation of the NATtrol™ master stock used in the manufacture of this product. Samples were inoculated into appropriate cells and incubated for appropriate time and temperature. The inoculated cell cultures were observed for signs of cytopathic effects (CPE). NATtrol™ samples must demonstrate no CPE to verify inactivation.

**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.**

### Statement of Conformance:

The results reported in this Certificate of Analysis comply with the requirements of ZeptoMetrix's Quality Management System and specifications described in the Master Production Record.